



Standard Specification for Rubber Surgical Gloves¹

This standard is issued under the fixed designation D 3577; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense.

^{ε1} NOTE—An annex was deleted and the information was included in the Performance Requirements section editorially in March 2002.

^{ε2} NOTE—Sections 7.1.5 and 7.1.7 were revised editorially to correct measurement units in April 2002.

1. Scope

1.1 This specification describes certain requirements for packaged sterile rubber surgical gloves used in conducting surgical procedures.

1.2 The following safety hazards caveat pertains only to the test method portion, Section 8, of this specification: *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:

D 412 Test Methods for Vulcanized Rubber and Thermoplastic Elastomers—Tension²

D 573 Test Method for Rubber—Deterioration in an Air Oven²

D 3767 Practice for Rubber—Measurement of Dimensions²

D 5151 Test Method for Detection of Holes in Medical Gloves³

D 5712 Test Method for the Analysis of Aqueous Extractable Protein in Natural Rubber and its Products Using the Modified Lowry Method³

D 6124 Test Method for Residual Powder on Medical Gloves³

D 6499 Test Method for the Immunological Measurement of Antigenic Protein in Natural Rubber and its Products³

2.2 Other Documents:

ISO 2859 Sampling Procedures and Tables for Inspection by Attributes⁴

U.S. Pharmacopeia⁵

3. Classification

3.1 *Type 1*—Gloves compounded primarily from natural rubber latex.

3.2 *Type 2*—Gloves compounded from a rubber cement or from synthetic rubber latex.

4. Materials and Manufacture

4.1 Any rubber polymer compound that permits the glove to meet the requirements of this specification.

4.2 A lubricant that meets the current requirements of the U.S. Pharmacopeia for Absorbable Dusting Powder may be applied to the glove. Other lubricants may be used if their safety and efficacy have been previously established.

4.3 The inside and outside surface of the rubber surgical gloves shall be free of talc.

5. Significance and Use

5.1 The specification is intended as a reference to the performance and safety of rubber surgical gloves. The safe and proper use of rubber surgical gloves is beyond the scope of this specification.

6. Sampling

6.1 For referee purposes, gloves shall be sampled and inspected in accordance with ISO 2859. The inspection levels and acceptable quality levels (AQL) shall conform to those specified in Table 1, or as agreed between the purchaser and the seller, if the latter is more comprehensive.

7. Performance Requirements

7.1 Gloves, sampled in accordance with Section 6, shall meet the following referee performance requirements:

7.1.1 Comply with requirements for sterility when tested in accordance with 8.2.

7.1.2 Be free from holes when tested in accordance with 8.3.

¹ This specification is under the jurisdiction of ASTM Committee D11 on Rubber and is the direct responsibility of Subcommittee D11.40 on Consumer Rubber Products.

Current edition approved Nov. 10, 2001. Published January 2002. Originally published as D 3577 – 77. Last previous edition D 3577 – 01^{ε1}.

² Annual Book of ASTM Standards, Vol 09.01.

³ Annual Book of ASTM Standards, Vol 09.02.

⁴ Available from American National Standards Institute, 25 West 43rd St., 4th Floor, New York, NY 10036.

⁵ U.S. Pharmacopeia, latest edition, Mack Publishing Co., Easton, PA 19175.

TABLE 1 Performance Requirements

Characteristic	Related Defects	Inspection Level	AQL
Sterility	fails sterility	^A	N/A
Freedom from holes	holes	I	1.5
Physical dimensions	length, width, and thickness	S-2	4.0
Physical properties	before aging, after accelerated aging	S-2	4.0
Powder Free Residue	Exceeds Maximum Limit	N=5	N/A
Protein Content	Exceeds Recommended Maximum Limit	N=3	N/A
Powder Amount	Exceeds Recommended Maximum Limit	N=2	N/A
Antigenic Protein Content	Exceeds Recommended Maximum Limit	N=1	N/A

^A See U.S. Pharmacopeia.

7.1.3 Have consistent physical dimensions in accordance with 8.4.

7.1.4 Have acceptable physical property characteristics in accordance with 8.5.

7.1.5 Have a powder residue limit of 2.0 mg in accordance with 8.6.

7.1.6 Have a recommended aqueous soluble protein content limit of 200 µg/dm² in accordance with 8.7 and Annex A1 or have a recommended antigenic protein content limit of 10 µg/dm² in accordance with 8.9 and Annex A2.

7.1.7 Have a recommended maximum powder limit of 15 mg/dm² in accordance with 8.8.

8. Referee Test Methods

8.1 The following tests shall be conducted to assure the requirements of Section 7 as prescribed in Table 1:

8.2 *Sterility Test*—Testing for sterility shall be conducted in accordance with the latest edition of the U.S. Pharmacopeia.

8.3 *Freedom from Holes*—Testing for freedom from holes shall be conducted in accordance with Test Method D 5151.

8.4 Physical Dimensions Test:

8.4.1 The gloves shall comply with the dimension requirements specified in Table 2.

8.4.2 The length shall be expressed in millimetres as measured from the tip of the second finger to the outside edge of the cuff.

8.4.3 The width of the palm shall be expressed in millimetres as measured at a level between the base of the index finger and the base of the thumb. Values of width per size other than listed shall meet the stated tolerance specified in Table 2.

8.4.4 The minimum thickness shall be expressed in millimetres as specified in Table 2 when using a dial micrometer

described in Practice D 3767 and in the locations indicated on Fig. 1. For referee tests, cutting the glove is necessary to obtain single-thickness measurements.

8.4.5 *Precision and Bias*—The precision and bias of measuring glove dimensions are as specified in Practice D 3767.

8.5 Physical Requirements Test:

8.5.1 Before and after accelerated aging, the gloves shall conform to the physical requirements specified in Table 3. Tests shall be conducted in accordance with Test Methods D 412.

8.5.2 Accelerated aging tests shall be conducted in accordance with Test Method D 573. Test the gloves by either one of the following methods:

8.5.2.1 After being subjected to a temperature of 70 ± 2°C for 166 ± 2 h, the tensile strength and ultimate elongation shall not be less than the values specified in Table 3. This method shall be the condition for referee tests.

8.5.2.2 After being subjected to a temperature of 100 ± 2°C for 22 ± 0.3 h, the tensile strength and ultimate elongation shall not be less than the values specified in Table 3.

8.5.3 *Precision and Bias*—The precision and bias of determining tensile strength and ultimate elongation of gloves are as specified in Test Methods D 412.

8.6 Powder Free Gloves:

8.6.1 Determine the powder residue using Test Method D 6124.

8.7 Aqueous Extractable Protein Content:

8.7.1 Determine the aqueous extractable protein (µg/mL) using Test Method D 5712 for each glove sample tested.

8.7.2 Determine the total µg of aqueous extractable protein in each glove sample by multiplying the result from 8.7.1 by the total volume of extractant used for that specific glove sample. If the glove sample is less than a whole glove, then adjust the protein results to reflect the amount of protein in the whole glove.

8.7.3 Determine the square decimeters for the glove size. Multiply the minimum length and nominal width found in Table 2 and convert to dm² using (dm²/mm²) (mm²/10 000) (4). Four (4) is the factor for all inside and outside surface areas.

8.7.4 Determine the aqueous extractable protein content of a glove sample by dividing the result from 8.7.2 (total µg of protein) by 8.7.3 (total surface area of glove).

8.7.5 If the sample is more than one (1) glove, use the average µg/dm² of protein for the number of gloves tested in the sample.

8.8 Powdered Gloves:

TABLE 2 Dimensions and Tolerances

Designation	Size								Tolerance
	5½	6	6½	7	7½	8	8½	9	
Length, mm	245	265	265	265	265	265	265	265	min
Width, mm	70	76	83	89	95	102	108	114	±6
Thickness, mm:									
Finger					0.10				min
Palm					0.10				min
Cuff					0.10				min




FIG. 1 Location of Thickness Measurements

TABLE 3 Physical Requirements

Type	Tensile Strength	Before Aging		After Accelerated Aging	
		Ultimate Elongation	Stress at 500 % Elongation	Tensile Strength	Ultimate Elongation
I	24 MPa, min	750 % min	5.5 MPa, max	18 MPa, min	560 % min
II	17 MPa, min	650 % min	7.0 MPa, max	12 MPa, min	490 % min

8.8.1 Determine the recommended maximum powder limit using Test Method D 6124 for powdered gloves.

8.8.2 Determine the square decimeters for the glove size as in 8.7.3.

8.9 Antigenic Protein Content:

8.9.1 Determine the extractable antigenic protein (µg/mL) using Test Method D 6499 for each glove sample tested.

8.9.2 Determine the total microgram of extractable antigenic protein in each glove sample by multiplying the result from 8.9.1 by the total volume of extractant used for that specific glove sample.

8.9.3 Determine the square decimeter for the glove size as in 8.7.3.

8.9.4 Determine the extractable antigenic protein content of a glove sample by dividing the result from 8.9.2 (total microgram of antigenic protein) by 8.9.3 (total surface area of glove).

9. Acceptance

9.1 Gloves will be considered to meet the referee perfor-

mance requirements when test results conform to the requirements prescribed in Table 1.

9.2 Retests or reinspections are permissible under the provisions of the U.S. Pharmacopeia and ISO 2859.

10. Packaging and Package Marking

10.1 *Packaging*—Packaging shall be provided to maintain sterility after sterilization during shipping and storage and permit opening without contamination of the gloves.

10.2 Marking:

10.2.1 Gloves shall have an appropriate marking or be color-coded to designate size.

10.2.2 Inner wrappers or wallets, if used, shall bear a size marking to be located on the outside of the wallet or wrapper.

10.2.3 Packages shall bear markings for the contents to include the glove size, instructions for opening, the legend “sterile,” and a manufacturing lot number.

10.2.4 The outermost case shall be labeled on one or more end panels with the glove size, the legend “sterile,” and a manufacturing lot number.

10.2.5 All levels of packaging shall conform to all appropriate government labeling regulations.

11. Keywords

11.1 surgical; gloves; rubber

ANNEXES

(Mandatory Information)

A1. PROTEIN CONTENT

A1.1 The current assay precision is large enough that only a recommended limit can be considered.

A1.2 Consideration should be given to the relative repeatability and reproducibility when reporting test method results.

A1.3 Reasonable allowance should be given for test results in excess of the recommended limit until greater precision of the method can be attained.

A2. ANTIGENIC PROTEIN CONTENT

A2.1 The current assay precision is large enough that only a recommended limit can be considered.

A2.2 Consideration should be given to the relative repeatability and reproducibility when reporting test method results.

A2.3 A pooled sample from three individual NR specimens or products as extracted in Test Method D 5712 is permitted for use as the extraction sample.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).