




**Unjointed/Rigid IEC Finger Probe**

**Model UFP-01**



The Unjointed/rigid Finger Probe is made for forceful insertion where the flexibility of the Jointed Finger Probe would be a hindrance. The UFP-01 meets IEC 60950 and 61010, CSA and UL requirements. This probe is precision made of flash chrome plated steel and the rest of the instrument of Delrin®. The handle can be threaded for use with our digital force gauges.

All ED&D probes come with NIST traceable certificates of calibration. There is an additional fee for the data, and it must be ordered in advance. ED&D operates an ISO/IEC 17025 calibration lab.